



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/634,410	08/05/2003	Jorge A. Groiso	P/1825-11	7948

7590 09/20/2007
OSTROLENK FABER GERB & SOFFEN
1180 AVENUE OF THE AMERICAS
NEW YORK, NY 10036-8403

EXAMINER

ARAJ, MICHAEL J

ART UNIT	PAPER NUMBER
----------	--------------

3733

MAIL DATE	DELIVERY MODE
-----------	---------------

09/20/2007

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/634,410

Applicant(s)

GROISO, JORGE A.

Examiner

Michael J. Araj

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 13 September 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-31 is/are pending in the application.
- 4a) Of the above claim(s) 6-20 and 23-31 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-5, 21 and 22 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on September 13, 2007 has been entered.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-3, 5, 21 and 22 are rejected under 35 U.S.C. 102(e) as being anticipated by Kanner et al. (U.S. Patent No. 6,767,356).

Kanner et al. disclose a clip that has at least two straight engagement legs (1102, 1104), extending parallel to one another including respective distal bridging tips and respective proximal insertion tips (1102A, 1102B, 1104A and 1104B) as well as a connecting bridge coupled to the distal bridging tips of the two engagement legs that are

at least two elongated sections (1110) extending along side on another. The elongated section extends along a non-linear trajectory and includes one of a depression and a dome (see Fig. 67). Also disclosed is that the engagement legs and the connecting bridges are constructed from a biocompatible material (Col. 2, Paragraph 1).

Furthermore, the two elongated section includes respective bulges that together form a space of separation between elongated sections. Applying a force to the elongated sections of the clip in an area of the non-linear deformable region causes the proximal insertion tips to pivot which can be seen between figure 68-71. The force is selected from at least one to separate and to unite the two elongated sections. For further clarification the elongated section begins from the point of 1108 and ends at the other side of the clip of 1108 that can be seen in fig. 69. It can be seen that a force is being applied to this part of the elongated section which caused the insertion tips to pivot.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 4 is rejected under 35 U.S.C. 103(a) as being unpatentable over Kanner et al. (U.S. Patent No. 6,767,356) in view of Wevers et al. (U.S. Patent No. 4,456,006).

Kanner et al. disclose the claimed invention except for the engagement legs being provided with gripping surfaces. Wevers et al. disclose notches (14) designed to

Art Unit: 3733

facilitate gripping when implanted (col. 2 lines 10-13). It would have been obvious to one skilled in the art at the time the invention was made to have created the device of Kanner et al. with engagement legs having gripping surfaces in view of Wevers et al., in order to have a better grip of the implant.

Response to Arguments

Applicant's arguments filed September 13, 2007 have been fully considered but they are not persuasive. With respect to the arguments that Kanner is not suitable for repair of bone tissue and thus no mention of bone tissue is made in Kanner; even though it is not specifically disclosed that this device is used in bone tissue it is still possible to function with it.. Applicant is reminded that an anticipation under 35 U.S.C. 102(b) or 102(e) is established when a single prior art reference discloses, either expressly or under principles of inherency, each and every element of of a claimed invention. *RCA Corp. v. Applied Digital Data System, Inc.*, 730 F.2d 1440, 221 USPQ 385 (Fed. Cir. 1984). Furthermore, it is well settled that the law of anticipation does not require that the reference teach what appellants is teaching or has disclosed, but only that the claims on appeal "read on" something disclosed in the reference, i.e. all limitation of the claims are found in the reference. *Kalman v. Kimberly Clark Corp.*, 713 F.2d 760, 218 USPQ 781 (Fed. Cir. 1083). Moreover, it is not necessary for the applied reference to expressly disclose or describe a particular element or limitation of a rejected claim word for word as in the rejected claim so long as the reference inherently

discloses that element or limitation. *Standard Havens Products Inc. v. Gencor Industries Inc.*, 953 F.2d 1360, 21 USPQ 2d. 1321 (Fed. Cir. 1991).

Also, applicant claims that Kanner's device does not have at least two straight engagement legs extending parallel to one another. It is unclear to how the applicant is interpreting how fig 67 does not show this. The distal half of the legs are parallel to each other (1102 and 1104).

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael J. Araj whose telephone number is 571-272-5963. The examiner can normally be reached on M-F 8am-5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


MJA


EDUARDO C. ROBERT
SUPERVISORY PATENT EXAMINER